



MISSOURI DEPARTMENT OF HEALTH
STATE PUBLIC HEALTH LABORATORY
DATAMASTER MAINTENANCE REPORT

RECEIVED
DHSS Breath Alcohol Program
By Carol Day at 12:22 pm, Mar 09, 2010

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

DATAMASTER SN 204159 INV# 127276 DATE OF INSPECTION 03-02-10

LOCATION OF INSTRUMENT (STREET AND CITY) 415 THIRD PLATTE CITY MO TIME OF INSPECTION 2114

CHECKLIST: Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

☒ DIAGNOSTIC CHECK (PRINTOUT ATTACHED)

☒ COMPUTER

☒ DETECTOR

☒ PROGRAM

☒ FILTERS

☒ HEATERS SAMPLE CHAMBER 50 °C

☒ QUARTZ STANDARD

☒ FLOW DETECTOR

☒ CALIBRATION

☒ PUMP HIGH SPEED

☒ PRINTER

☒ INDICATOR LIGHTS

☒ TIME AND DATE

☒ SIMULATOR TEMPERATURE (34 °C ± 0.2°C)

☒ CALIBRATION CHECK -

Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

TEST 1 ☒ .098

TEST 2 ☒ .099

TEST 3 ☒ .100

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

☒ NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)

REFUSALS |

(0-.04) 5

(.05-.09) 1

(.10-.14) 3

(.15-.19) 2

(Over .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

INSTRUMENT IS OPERATING WITHIN D.H.S.S. SPECS

REPCO MARKETING INC., LOT# 09001, BOTTLE# 0138, 10 VAPOR, EXP 04-22-11

INSPECTING OFFICER

SIGNATURE

CPL. S.M. SALFRANK #886

PRINT NAME

CPL. S.M. SALFRANK #886

TYPE II PERMIT NUMBER/EXPIRATION DATE

920257 / 12-07-11

TELEPHONE NUMBER

816.622.0800

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 09001

EXPIRATION DATE: April 22, 2011 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 09001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1214 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is April 23, 2009.

The expiration date for this lot number is April 22, 2011 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

BAC DataMaster

Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATA MASTER SERIAL NUMBER 204159

03/02/10
21:14

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 50C
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
WARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~

Operator Signature CPL S.M. SALFANIK #886

BAC DataMaster

Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATA MASTER SERIAL NUMBER 204159

03/02/10

TESTING OFFICER:

SUPERVISOR S/M
OFFICER I.D.: 886
PERMIT NUMBER: 920257
EXPIRATION DATE: 12/07/11
MISCELLANEOUS DATA:
.10 VAPOR

--- SUPERVISOR MODE ---

TEST	VERIFIED	TIME
BLANK TEST	.000	21:21
INTERNAL STANDARD		21:21
EXTERNAL STANDARD	.098	21:22
BLANK TEST	.000	21:22
EXTERNAL STANDARD	.099	21:23
BLANK TEST	.000	21:23
EXTERNAL STANDARD	.100	21:24
BLANK TEST	.000	21:24

N = 3
SIM. = .1
RMS. = .099

Operator Signature CPL S.M. SALFANIK #886

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204159
03/02/10

ARREST TIME: 00:00
SUBJECT NAME:
RFI/TEST
DOB: 11/11/11 SEX: M
STATE/D.L.: MO/111111111
ARRESTING OFFICER:
N/A
OFFICER I.D.: N/A
TESTING OFFICER:
SALFRANK/S/M
OFFICER I.D.: 886
PERMIT NUMBER: 920257
EXPIRATION DATE: 12/07/11
MISCELLANEOUS DATA:
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	21:28
INTERNAL STANDARD	VERIFIED	21:28
RADIO INTERFERENCE		

Operator Signature CPL. S.M. SALFRANK #886

Printed on recycled paper with agri-based inks

CMSU 2208-02

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



STEVEN SALFRANK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 12/07/09

Number 920257

Expires 12/07/2011

MO 580-0771 (7-88)

Interim Director

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)